



**DR. B. C. ROY COLLEGE OF PHARMACY AND ALLIED HEALTH SCIENCES, DURGAPUR**

Affiliated to MAKAUT, WB and WBSCT & VE & SD

Approved by PCI | NAAC & NBA (UG Pharmacy) Accredited | NIRF Ranked (2020-2024)

CAMPUS: Dr. Meghnad Saha Sarani, Bidhannagar, Durgapur-713206, Dist.-Paschim Burdwan (W.B), India  
☎ 0343-2532679/0343-2532678, ✉ bcrpc\_dgp@yahoo.co.in

BCRCP/INQ/03/26/028

Date - 10.03.2026

**Notice for Submission of Quotation**

Quotations are invited within 10 days of publication of this notice for the following Instrument. Quotation is to be sent to [purchase@bcrpc.org](mailto:purchase@bcrpc.org) in pdf as well as in excel format. Quotations must include GSTIN., PAN/ TAN and Trade Licence of the Vendor/ Supplier. User Certificate may also be sought before confirming the order.

Sl. No	Name of the Instrument	Required Quantity	Rate
1	<p>Bench-top / Laboratory Lyophilizer for Pharmaceutics Research Lab (PG) PGRL - 1</p> <p><b>Parameters</b></p> <p>Shelf area (MTP) - 0.08 m<sup>2</sup>, Condenser temp - Up to - 85 °C</p> <p>Trap capacity Max 5 L, Ice capture (24 h) - 4-6 kg</p> <p>Pump rate 2 L/s, Ultimate vacuum &gt;= 1 Pa</p> <p>Shelf Heating Preferable, Manifold - Yes</p> <p>Power supply - 250 V, 50 Hz</p> <p>Kindly include the following details in your quotation:</p> <ul style="list-style-type: none"><li>• Complete model name and technical specifications</li><li>• Price breakup (basic unit, accessories, vacuum pump, manifold/shelves)</li><li>• Warranty period</li><li>• AMC details after warranty</li><li>• Delivery time</li><li>• Installation and training details</li><li>• Service support availability in Kolkata / West Bengal</li><li>• Details of your local service engineer or authorized service center in Kolkata</li><li>• Product brochure / datasheet</li><li>• List of academic or pharma institute users (if available)</li></ul>	1	

(Prof. Samir Kumar Samanta)  
Principal

Prof. (Dr.) Samir Kumar Samanta  
M. Pharm., Ph.D (J.U.)  
Principal  
Dr. B. C. Roy College of Pharmacy & AHS  
Durgapur, West Bengal-713206

