



DR. B. C. ROY COLLEGE OF PHARMACY AND ALLIED HEALTH SCIENCES, DURGAPUR

Affiliated to MAKAUT, WB and WBSCT & VE & SD

Approved by PCI | NAAC & NBA (UG Pharmacy) Accredited | NIRF Ranked (2020-2024)

CAMPUS: Dr. Meghnad Saha Sarani, Bidhannagar, Durgapur-713206, Dist.-Paschim Burdwan (W.B), India

☎ 0343-2532679/0343-2532678, ✉ bcrpc_dgp@yahoo.co.in

BCRCP/INQ/12/25/027

Date: 04-12-2025

Notice for submission of Quotation

Quotation are invited within 10 days of publication of notice for the following Instruments for Dip. UG & PG Laboratories. Quotation are to be sent purchase@bcrpc.org in pdf as well in excel format. Quotation must include GSTIN, PAN/TAN & Trade Licence of the Vender / Supplier.

Sl. No.	Name of the Instruments	SPECIFICATIONS	Req. Quantity	Unit	Mfg By	Rate/unit	GST%
1	Analytical balance	Capacity 200–300 g; Readability 0.1 mg; Internal calibration; LCD display; Draft shield	1	Pcs			
2	Andreasen's Pipette	25–50 ml Borosilicate; complete with sedimentation bottle and stand	5	Pcs			
3	Clarity test apparatus	3–6 place; LED light; adjustable sample holders; suitable for IP/BP clarity tests	1	Pcs			
4	Compound microscope	40x–1000x; LED illumination; mechanical stage; fine/coarse focus	5	Pcs			
5	Digital Balance-0.1mg sensitivity	0.1 mg readability; 120–200 g capacity; internal calibration; draft shield	1	Pcs			
6	Digital Colorimeter	400–700 nm; LED filters; accuracy ±2%; cuvette set	2	Pcs			
7	Digital Melting point apparatus	Ambient–350°C; accuracy ±1°C; digital controller; capillary holder	2	Pcs			
8	Digital Weight machine 10 mg sen.	Readability 0.01 g; capacity 5–10 kg; LCD display; tare function	7	Pcs			
9	Glass Distillation Unit	Borosilicate; 1–2 L/hr output; quartz heater; condenser + boiler assembly	1	Pcs			



Prof. (Dr.) Samir Kumar Samanta
M. Pharm., Ph.D (J.U.)
Principal

Dr. B. C. Roy College of Pharmacy & AHS
Durgapur, West Bengal-713206



DR. B. C. ROY COLLEGE OF PHARMACY AND ALLIED HEALTH SCIENCES, DURGAPUR

Affiliated to MAKAUT, WB and WBSCT & VE & SD

Approved by PCI | NAAC & NBA (UG Pharmacy) Accredited | NIRF Ranked (2020-2024)

CAMPUS: Dr. Meghnad Saha Sarani, Bidhannagar, Durgapur-713206, Dist.-Paschim Burdwan (W.B), India
☎ 0343-2532679/0343-2532678, ✉ bcrpc_dgp@yahoo.co.in

Sl. No.	Name of the Instruments	SPECIFICATIONS	Req. Quantity	Unit	Mfg By	Rate/unit	GST%
10	Electric water bath unit	6-12 holes; ambient-100°C; accuracy ±1°C; SS chamber; thermostat	1	Pcs			
11	Glass Distillation unit with complet set(including heating Mantel,5 ltr. Cap. R B-5 ltr,reflux condenser,colletcor etc. (Borosilicate Glass)	Borosilicate; 5 L mantle; RB flask 5 L; reflux condenser; full assembly	5	Pcs			
12	Height and weight chart	Laminated; 0-200 cm height + weight chart	1	pcs			
13	Hot air oven (95L)	95 L; ambient-250°C; digital temp control; SS chamber; air circulation	1	Pcs			
14	Hot plate	SS plate; 300°C max; 500-1000W; thermostat control	3	Pcs			
15	Full Human skeleton	Life-size 170-180 cm; PVC; removable limbs; stand included	1	Pcs			
16	MAGNETIC STIRRER 1ML (without heater)	1 L capacity; 100-2000 rpm; chemical-resistant top	5	Pcs	REMI		
17	MAGNETIC STIRRER 1MLH	1 L capacity; heater up to 300°C; 100-2000 rpm	5	Pcs	REMI		
18	Magnetic stirrer with thermostat	100-2000 rpm; heating 50-300°C; digital thermostat	10	Pcs			
19	Malique water system	18.2 MΩ-cm ultrapure water; TOC <10 ppb; UV + UF purification	1	Pcs			
20	Simple Microscope	10x-30x magnification; focusing stand; LED light	19	Pcs			
21	Microscope with camera	40x-1000x; LED; mechanical stage; 2-5 MP USB camera	1	Pcs			
22	Nessler's cylinder 50 ml (Borosilicate)	50 ml borosilicate; matched pair; graduation marks	20	Pcs			



Prof. (Dr.) Samir Kumar Samanta
M. Pharm., Ph.D (J.U.)

Principal

Dr. B. C. Roy College of Pharmacy & AHS
Durgapur, West Bengal-713206



DR. B. C. ROY COLLEGE OF PHARMACY AND ALLIED HEALTH SCIENCES, DURGAPUR

Affiliated to MAKAUT, WB and WBSCT & VE & SD

Approved by PCI | NAAC & NBA (UG Pharmacy) Accredited | NIRF Ranked (2020-2024)

CAMPUS: Dr. Meghnad Saha Sarani, Bidhannagar, Durgapur-713206, Dist -Paschim Burdwan (W.B), India

☎ 0343-2532679/0343-2532678, ✉ bcrpc_dgp@yahoo.co.in

Sl. No.	Name of the Instruments	SPECIFICATIONS	Req. Quantity	Unit	Mfg By	Rate/unit	GST%
23	Ointment spatula	SS spatula; 6-8 inch; flexible blade; ergonomic handle	20	Pcs			
24	Sahlis hemoglobinometer	Complete set; comparator; graduated Hb tube; stirrer	10	Pcs			
25	Snellen's chart	Laminated; 6 m distance; English letters	1	Pcs			
26	Standard sieves, sieve no. 8, 10, 12, 22, 24, 44, 54, 60, 80, 85, 100, 120-4 inch dia Brass	Brass sieves; 4-inch diameter; mesh 8-120; IS:460 compliant	12	Pcs			
27	Tablet disintegration test apparatus IP	1-2 basket; 30 cycles/min; 37°C ±0.5°C; digital timer	1	Pcs			
28	Tablet dissolution test apparatus IP (Digital single Unit)	USP I/II; 1-3 vessels; 25-200 rpm; digital display; PID control	1	Pcs			
29	Thiele's tube (Borosilicate)	Borosilicate; 150-200 mm; uniform heating design	16	Pcs			
30	Tissue slides set Permanent with box	50/100 slides; labeled; storage box included	25	Pcs			
31	Tonometer	Schiötz type; plunger + weights; complete with case	1	Pcs			
32	Vacuum desiccator (250 mm)	250 mm; borosilicate/polycarbonate; vacuum valve; porcelain plate	2	Pcs	Tarson		
33	Variable volume Micropipette (10-200 µl)	Adjustable 10-200 µl; accuracy ±1%; autoclavable	1	Pcs			
34	Variable volume Micropipette (1-20 µl)	Adjustable 1-20 µl; accuracy ±1%; autoclavable	1	Pcs			
35	Variable volume Micropipette (200-1000 µl)	Adjustable 200-1000 µl; accuracy ±1%; autoclavable	1	Pcs			
36	Wall mounted distillation unit SS	SS body; 1-2 L/hr; quartz heater; automatic cutoff	2	Pcs			



Prof. (Dr.) Samir Kumar Samanta
M. Pharm., Ph.D (J.U.)

Principal

Dr. B. C. Roy College of Pharmacy & AHS
Durgapur, West Bengal-713206



DR. B. C. ROY COLLEGE OF PHARMACY AND ALLIED HEALTH SCIENCES, DURGAPUR

Affiliated to MAKAUT, WB and WBSCT & VE & SD

Approved by PCI | NAAC & NBA (UG Pharmacy) Accredited | NIRF Ranked (2020-2024)

CAMPUS: Dr. Meghnad Saha Sarani, Bidhannagar, Durgapur-713206, Dist -Paschim Burdwan (W.B), India

☎ 0343-2532679/0343-2532678, ✉ bcrpc_dgp@yahoo.co.in

Sl. No.	Name of the Instruments	SPECIFICATIONS	Req. Quantity	Unit	Mfg By	Rate/unit	GST%
37	Westergren apparatus with tubes (ESR STAND WITH 6 PIPETTE)	6-tube ESR stand; 0-200 mm Westergren pipettes; leveling base	1	Pcs			

(Prof. Samir Kumar Samanta)
Principal

Prof. (Dr.) Samir Kumar Samanta
M. Pharm., Ph.D (J.U.)
Principal
Dr. B. C. Roy College of Pharmacy & AHS
Durgapur, West Bengal-713206



[Click here for Excel File](#)